



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Wayac Scales & Calibration, Inc.***  
***2899 Hilliard Rome Road, Hilliard, OH 43026***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Dimensional, Mass, Force, Weighing, Time & Frequency, Mechanical, Thermodynamic, and Electrical Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

April 8, 2003

*Issue Date:*

November 2, 2017

*Expiration Date:*

November 2, 2019

*Accreditation No.:*

59301

*Certificate No.:*

L17-478

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

| MEASURED INSTRUMENT, QUANTITY OR GAUGE             | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE            | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|--|--|--|
| Calipers <sup>FO</sup>                             | 1.27 mm to 914.4 mm<br>(0.05 in to 36 in)              | (2.82 + .127L) $\mu$ m<br>[(111 + 5L) $\mu$ in]                              | Gage Blocks<br>Length Standards<br>Ring Gage       |
| Height Gages <sup>FO</sup>                         | 1.27 mm to 914.4 mm<br>(0.05 in to 36 in)              | (11.3 + .0254L) $\mu$ m<br>[(445 + 1L) $\mu$ in]                             | Gage Blocks<br>Length Standards                    |
| Outside Micrometers <sup>FO</sup>                  | 1.27 mm to 914.4 mm<br>(0.05 in to 36 in)              | (1.57+ .0762L) $\mu$ m<br>[(62 + 3L) $\mu$ in]                               | Gage Blocks<br>Length Standards                    |
| Depth Micrometers <sup>FO</sup>                    | 1.27 mm to 457.2 mm<br>(0.05 in to 12 in)              | (1.78 + .152L) $\mu$ m<br>[(70 + 6L) $\mu$ in]                               | Gage Blocks<br>Length Standards                    |
| Inside Micrometers <sup>FO</sup>                   | 1.27 mm to 914.4 mm<br>(0.05 in to 36 in)              | (1.14 + .102L) $\mu$ m<br>[(45 + 4L) $\mu$ in]                               | Gage Blocks  |
| Indicators <sup>FO</sup>                           | 1.27 mm to 50.8 mm<br>(0.05 in to 2 in)                | (.763 + 153L) $\mu$ m<br>[(30 + 6L) $\mu$ in]                                | Indicator Calibrator                               |
| Pin Gages <sup>F</sup>                             | 0.279 4 mm to 25.4 mm<br>(0.011 in to 1 in)            | 0.000 86 mm<br>(0.000 034 in)  | Super-Micrometer with<br>Grade 1 Gage Blocks       |
| Smooth Plug Gages <sup>F</sup>                     | 0.279 4 mm to 127 mm<br>(0.011 in to 5 in)             | 0.000 86 mm<br>(0.000 034 in)  | Super-Micrometer with<br>Grade 1 Gage Blocks       |
| Thread Plug Gages / Pitch Diameter <sup>F</sup>    | 1.0 x 0.2 to 127 x 10<br>(0-80 to 5-5 inches)          | 0.003 8 mm<br>(0.001 5 in)   | Super-Micrometer with<br>Thread Wires              |
| Smooth Ring Gages <sup>F</sup>                     | 3.175 mm to 76.2 mm<br>(0.012 5 in to 3 in)            | 0.000 48 mm<br>(0.000 019 in)  | Super-Micrometer with<br>Grade 1 Gage Blocks       |
| Rules & Tapes <sup>FO</sup><br>Error of Indication | 9 144 mm maximum<br>(360 in maximum)                   | 0.39 mm<br>(0.016 in)  | Master Ruler<br>Gage Blocks                        |
| Surface Plates <sup>FO</sup><br>Repeat Reading     | 0.000 26 mm to 0.508<br>mm<br>(0.000 01 in to 0.02 in) | 0.001 3 mm<br>(0.000 05 in)  | Repeat-O-Meter                                     |
| Surface Plates <sup>FO</sup><br>Flatness           | 152.4 mm to 1 219.2 mm<br>(6 in to 48 in)              | 0.000 76 mm<br>(0.000 03 in)   | Planekator<br>Grade AA 24"                         |
| Protractors <sup>FO</sup>                          | Up to 90°  | 0.20°  | Gage Blocks, Surface<br>Plate and Sine Bar         |

### Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE          | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Equipment to Source<br>DC Voltage <sup>FO</sup> | 0.2 V to 2.00 V                             | 0.003 % of reading + 7 $\mu$ V   | Transmille 3041A                                   |
|   | 2.1 V to 20.0 V                             | 0.002 5 % of reading + 480 $\mu$ V   |  |
|   | 21 V to 200 V                               | 0.003 % of reading + 4 300 $\mu$ V   |  |
|   | 201 V to 1 000 V                            | 0.003 % of reading + 4 800 $\mu$ V   |  |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                 | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED |
|---|---|---|---|
| AC Voltage- Source <sup>FO</sup><br>at listed Frequencies |   |   | Transmille 3041A  |
| 10 Hz to 44 Hz  | 20 mV to 202 mV                                   | 0.2 % of reading + 51 $\mu$ V   |   |
| 45 Hz to 999 Hz   | 20 mV to 202 mV                                   | 0.04 % of reading + 88 $\mu$ V  |   |
| 1 kHz to 19.999 kHz                                       | 20 mV to 202 mV                                   | 0.09 % of reading + 38 $\mu$ V  |   |
| 20 kHz to 99.999 kHz                                      | 20 mV to 202 mV                                   | 0.3 % of reading + 71 $\mu$ V   |   |
| 100 kHz to 500 kHz  | 20 mV to 202 mV                                   | 0.8 % of reading + 1 281 $\mu$ V  |   |
| AC Voltage- Source <sup>FO</sup><br>at listed Frequencies |   |   |   |
| 10 Hz to 44Hz   | 0.2 V to 2.02 V                                   | 0.2 % of reading + 363 $\mu$ V  |   |
| 45 Hz to 999 Hz   | 0.2 V to 2.02 V                                   | 0.04 % of reading + 152 $\mu$ V   |   |
| 1 kHz to 19.999 kHz                                       | 0.2 V to 2.02 V                                   | 0.09 % of reading + 194 $\mu$ V   |   |
| 20 kHz to 99.999 kHz                                      | 0.2 V to 2.02 V                                   | 0.25 % of reading + 2 303 $\mu$ V   |   |
| 100 kHz to 500 kHz  | 0.2 V to 2.02 V                                   | 0.45 % of reading + 4 032 $\mu$ V   |   |
| AC Voltage- Source <sup>FO</sup><br>at listed Frequencies |   |   |   |
| 10 Hz to 44 Hz  | 2 V to 20.2 V                                     | 0.2 % of reading + 7 000 $\mu$ V  |   |
| 45Hz to 999 Hz  | 2 V to 20.2 V                                     | 0.035 % of reading + 2 921 $\mu$ V  |   |
| 1 kHz to 19.999 kHz                                       | 2 V to 20.2 V                                     | 0.07 % of reading + 1 463 $\mu$ V   |   |
| 20 kHz to 100 kHz   | 2 V to 20.2 V                                     | 0.22 % of reading + 37 300 $\mu$ V  |   |
| AC Voltage- Source <sup>FO</sup><br>at listed Frequencies |   |   | Transmille 3041A  |
| 30 Hz to 44 Hz  | 20 V to 202 V                                     | 0.06 % of reading + 52 mV   |   |
| 45 Hz to 999 Hz   | 20 V to 202 V                                     | 0.04 % of reading + 53 mV   |   |
| 1 kHz to 20 kHz   | 20 V to 202 V                                     | 0.09 % of reading + 257 mV  |   |
| AC Voltage- Source <sup>FO</sup><br>at listed Frequencies |   |   |   |
| 30 Hz to 45 Hz  | 200 V to 1 020 V                                  | 0.06 % of reading + 256 mV  |   |
| 45 Hz to 999 Hz   | 200 V to 1020 V                                   | 0.04 % of reading + 163 mV  |   |
| 1 kHz to 10 kHz   | 200 V to 1 020 V                                  | 0.15 % of reading + 472 mV  |   |
| DC Current - Source <sup>FO</sup>                         |   |   |   |
|   | 20 $\mu$ A to 202 $\mu$ A                         | 0.01 % of reading + 21 $\mu$ A  |   |
|   | 0.2 mA to 2.02 mA                                 | 0.008 % of reading + 29 $\mu$ A   |   |
|   | 2 mA to 20.2 mA                                   | 0.005 % of reading + 90 $\mu$ A   |   |
|   | 20 mA to 202 mA                                   | 0.008 % of reading + 1 281 $\mu$ A  |   |
|   | 0.2 A to 2.02 A                                   | 0.015 % of reading + 1 005 $\mu$ A  |   |
|   | 2 A to 30 A                                       | 0.04 % of reading + 2 660 $\mu$ A   |   |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                    | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED |
|--|---|---|---|
| AC Current – Source <sup>FO</sup><br>at the listed Frequency |   |   | Transmille 3041A  |
| 10 Hz to 44 Hz   | 20 $\mu$ A to 202 $\mu$ A                         | 0.2 % of reading + 60 $\mu$ A   |   |
| 45 Hz to 999 Hz  | 20 $\mu$ A to 202 $\mu$ A                         | 0.07 % of reading + 59 $\mu$ A  |   |
| 1 kHz to 10 kHz  | 20 $\mu$ A to 202 $\mu$ A                         | 0.8 % of reading + 62 $\mu$ A   |   |
| 10 Hz to 44 Hz   | 0.2 mA to 2.02 mA                                 | 0.2 % of reading + 53 $\mu$ A   |   |
| AC Current – Source <sup>FO</sup><br>at the listed Frequency |   |   |   |
| 45 Hz to 999 Hz  | 0.2 mA to 2.02 mA                                 | 0.06 % of reading + 33 $\mu$ A  |   |
| 1 kHz to 10kHz   | 0.2 mA to 2.02 mA                                 | 0.7 % of reading + 15 $\mu$ A   |   |
| 10 Hz to 44 Hz   | 2 mA to 20.2 mA                                   | 0.2 % of reading + 48 $\mu$ A   |   |
| 45 Hz to 999 Hz  | 2 mA to 20.2 mA                                   | 0.06 % of reading + 28 $\mu$ A  |   |
| 1 kHz to 10 kHz  | 2 mA to 20.2 mA                                   | 0.5 % of reading + 30 $\mu$ A   |   |
| AC Current – Source <sup>FO</sup><br>at the listed Frequency |   |   |   |
| 10 Hz to 44 Hz   | 20 mA to 202 mA                                   | 0.2 % of reading + 510 $\mu$ A  |   |
| 45 Hz to 999 Hz  | 20 mA to 202 mA                                   | 0.06 % of reading + 330 $\mu$ A   |   |
| 1 kHz to 5 kHz   | 20 mA to 202 mA                                   | 0.6 % of reading + 1 250 $\mu$ A  |   |
| AC Current – Source <sup>FO</sup><br>at the listed Frequency |   |   |   |
| 30 Hz to 44Hz  | 2 A to 30 A                                       | 0.2 % of reading + 62 400 $\mu$ A   |   |
| 45 Hz to 99 Hz   | 2 A to 30 A                                       | 0.09 % of reading + 31 900 $\mu$ A  |   |
| 100 Hz to 5 kHz  | 2 A to 30 A                                       | 0.3 % of reading + 4 000 $\mu$ A  |   |
| Equipment to Source<br>Resistance <sup>FO</sup>              |   |   | Transmille 3041A  |
|  | 0 to 100 $\Omega$                                 | 0.005 % of reading + 0.0052 $\Omega$  |   |
|  | 100 $\Omega$ to 1 k $\Omega$                      | 0.004 % of reading + 0.0407 $\Omega$  |   |
|  | 1 k $\Omega$ to 10 k $\Omega$                     | 0.004 % of reading + 0.407 $\Omega$   |   |
|  | 10 k $\Omega$ to 100 k $\Omega$                   | 0.004 % of reading + 4.075 $\Omega$   |   |
|  | 100 k $\Omega$ to 1 M $\Omega$                    | 0.01 % of reading + 121 $\Omega$  |   |
|  | 1 M $\Omega$ to 10 M $\Omega$                     | 0.035 % of reading + 449 $\Omega$   |   |
|  | 10 M $\Omega$ to 100 M $\Omega$                   | 0.5 % of reading + 6 850 $\Omega$   |   |
| Frequency - Source <sup>FO</sup>                             | 10 Hz to 500 kHz                                  | 0.4 % of reading  |   |
| DC Voltage Measure <sup>FO</sup>                             | 100 mV to 1 000 V                                 | 0.014 % of reading  | HP 34401A   |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT,<br>QUANTITY OR GAUGE                   | RANGE OR NOMINAL<br>DEVICE SIZE AS<br>APPROPRIATE | CALIBRATION AND<br>MEASUREMENT<br>CAPABILITY EXPRESSED<br>AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION<br>EQUIPMENT<br>AND REFERENCE<br>STANDARDS USED |
|---|---|---|---|
| AC Voltage Measure<br>at listed frequencies <sup>FO</sup>   |   |   | HP 34401A   |
| 5 Hz to 10 Hz   | 10 mV to 100 mV                                   | 0.048 % of reading  |   |
| 10 Hz to 20 kHz   | 10 mV to 100 mV                                   | 0.039 % of reading  |   |
| 20 kHz to 50 kHz  | 10 mV to 100 mV                                   | 0.041 % of reading  |   |
| 50 kHz to 100 kHz   | 10 mV to 100 mV                                   | 0.22 % of reading   |   |
| 100 kHz to 300 kHz  | 10 mV to 100 mV                                   | 0.5 % of reading  |   |
| 10 Hz   | 100 mV to 1 V                                     | 0.45 % of reading   |   |
| 20 kHz  | 100 mV to 1 V                                     | 0.177 % of reading  |   |
| AC Voltage Measure<br>at listed frequencies <sup>FO</sup>   |   |   | HP 34401A   |
| 50 kHz  | 100 mV to 1 V                                     | 0.188 % of reading  |   |
| 100 kHz   | 100 mV to 1 V                                     | 0.704 % of reading  |   |
| 50 Hz   | 1 V to 750 V                                      | 1.6 % of reading  |   |
| 10 kHz  | 1 V to 750 V                                      | 1.54 % of reading   |   |
| DC Current - Measure <sup>FO</sup>                          | up to 10 mA                                       | 1.3 % of reading  |   |
|   | 10 mA to 100 mA                                   | 0.78 % of reading   |   |
|   | 100 mA to 1.0 A                                   | 0.174 % of reading  |   |
|   | 1.0 A to 3.0 A                                    | 0.93 % of reading   |   |
| AC Current – Measure <sup>FO</sup><br>at listed frequencies |   |   |   |
| 10 Hz   | up to 1.0 A                                       | 0.81 % of reading   |   |
| 5 kHz   | up to 1.0 A                                       | 2.0 % of reading  |   |
| AC Current – Measure <sup>FO</sup><br>at listed frequencies |   |   |   |
| 10 Hz   | 1.0 A to 3.0 A                                    | 1.29 % of reading   |   |
| 5 kHz   | 1.0 A to 3.0 A                                    | 2.38 % of reading   |   |
| Resistance - Measure <sup>FO</sup>                          | up to 100.0 $\Omega$                              | 1.74 % of reading   | HP 34401A   |
|   | 100.0 $\Omega$ to 1.0 k $\Omega$                  | 0.014 % of reading  |   |
|   | 1.0 k $\Omega$ to 10.0 k $\Omega$                 | 0.14 % of reading   |   |
|   | 10.0 k $\Omega$ to 100.0 k $\Omega$               | 1.41 % of reading   |   |
|   | 100.0 k $\Omega$ to 1.0 M $\Omega$                | 0.027 1 % of reading  |   |
|   | 1.0 M $\Omega$ to 10.0 M $\Omega$                 | 1.44 % of reading   |   |
|   | 10.0 M $\Omega$ to 100.0 M $\Omega$               | 10.2 % of reading   |   |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE                          | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Frequency – Measure at Listed Voltages <sup>FO</sup>            |   |  | HP 34401A  |
| 100 mV  | 10 Hz                                       | 0.58 % of reading  |  |
| 100 mV  | 40 Hz                                       | 2.3% of reading  |  |
| 100 mV  | 300 kHz                                     | 0.43 % of reading  |  |
| Frequency – Measure at Listed Voltages <sup>FO</sup>            |   |  |  |
| 750 V   | 50 Hz                                       | 3.1 % of reading   |  |
| 750 V   | 10 kHz                                      | 16 % of reading  | Tektronix SG 503                                   |
| Oscilloscope – Flatness Relative to 50 kHz signal <sup>FO</sup> | 250 kHz to 100 MHz                          | 1.0 %  |  |
|   | 100 MHz to 250 MHz                          | 3.0 %  |  |
| Oscilloscope – Time Marks Horizontal Calibration <sup>FO</sup>  | 5 s to 1 ns                                 | (1 x 10 <sup>-7</sup> ) s  | Tektronix TG 501                                   |
| Oscilloscope – Band Width Vertical Calibration <sup>FO</sup>    | 0.2 V to 2.00 V                             | 0.003 % of reading + 7 $\mu$ V   | Transmille 3041A                                   |
|   | 2.1 V to 20.0 V                             | 0.002 5 % of reading + 480 $\mu$ V   |  |
|   | 21 V to 200 V                               | 0.003 % of reading + 4 300 $\mu$ V   |  |
|   | 201 V to 1 000 V                            | 0.003 % of reading + 4 800 $\mu$ V   |  |

### Mass, Force, and Weigh Devices

| MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE            | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|--|--|--|
| Weigh Balances <sup>FO</sup>           | 0.1 mg to 1 g  | 0.000 63g  | Class 1 Weights                                    |
|  | 1 g to 32 000 g  |  |  |
| Weigh Scales <sup>FO</sup>             | 226.796 g to 22 679.62 g (0.5 lb to 50 lb)             | 7.71g (0.017 lb)   | F Class Weights                                    |
|  | 45.36 g to 2 267.96 kg (100 lb to 5 000 lb)            | 1 723 g (3.8 lb)   |  |
|  | 2 721.55 kg to 18 143.69 kg (6 000 lb to 40 000 lb)    | 1 905 g (4.2 lb)   |  |
|  | 22 679.62 kg to 90 718.47 kg (50 000 lb to 200 000 lb) |  |  |
| Tension Compression <sup>FO</sup>      | 2.22 N to 4 448.222 N (0.5 lb to 1 000 lb)             | 10.68 N (2.4 lb)   | Indicator with NTEP Approved Load Cells            |
|  | 222.41 kN to 44 482 22 kN (50 lb to 10 000 lb)         | 0.0862 kN (19.38 lb)   | Indicator with NTEP Approved Load Cells            |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

| MEASURED INSTRUMENT, QUANTITY OR GAUGE   | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE                                | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|--|--|--|--|
| Indirect Verification of Rockwell Hardness Testers HRB <sup>FO</sup>   | HRB  | 2 HRB  | Certified Rockwell Test Blocks                     |
|  | 15 HRB to 31 HRB   |  |  |
|  | 31 HRB to 71 HRB   |  |  |
|  | 71 HRB to 100 HRB  |  |  |
| Electronic or Dial Pressure Gages, Commercial Grade, Medium Grade and Test Grade Transducers Stated values are gage pressure <sup>FO</sup> | 3.1 kPa to 2 068.4 kPa (0.45 psi to 300 psi)                               | 1.1 kPa (0.16 psi)   | Druck DPI104-300                                   |
|  | 2 075.32 kPa to 20 684.27 kPa (301 psi to 3 000 psi)                       | 10.48 kPa (1.6 psi)  | Druck DPI104-3000                                  |
|  | 20 691.2 kPa to 68 947.6 kPa (3 001 to 10 000 psi)                         | 38 kPa (5.5 psi)   | Druck DPI104-10000                                 |
| Magnahelics and Photohelics Gages <sup>FO</sup>  | 0.01 Pa to 249.088 Pa (0.13 in H <sub>2</sub> O to 1 in H <sub>2</sub> O)  | 3.74 Pa (0.015 in H <sub>2</sub> O)  | Manometers   |
|  | 0.03 Pa to 996.356 Pa (0.135 in H <sub>2</sub> O to 4 in H <sub>2</sub> O) | 11.21 Pa (0.045 in H <sub>2</sub> O)   |  |
| Vacuum <sup>FO</sup>   | -91.43 kPa to 0 kPa (-27 inHg to 0 inHg)                                   | 1.05 kPa (0.31 inHg)   | Druck DPI104-300                                   |
| Torque Wrenches, Torque Drivers <sup>FO</sup>  | 0.282 4 N·m to 2.824 N·m (40 ozf·in to 400 ozf·in)                         | 0.026 N·m (3.7 ozf·in)   | Torque Standards                                   |
|  | 0.564 9 N·m to 5.649 N·m (5 lbf·in to 50 lbf·in)                           | 0.19 N·m (1.7 lbf·in)  |  |
|  | 0.282 46 N·m to 28.246 N·m (25 lbf·in to 250 lbf·in)                       | 0.23 N·m (2.0 lbf·in)  |  |
|  | 11.298 5 N·m to 112.985 N·m (100 lbf·in to 1 000 lbf·in)                   | 0.80 N·m (7.1 lbf·in)  |  |
|  | 81.349 1 N·m to 813.491 N·m (720 lbf·in to 7 200 lbf·in)                   | 5.54 N·m (49 lbf·in)   |  |
| Durometers <sup>FO</sup><br>Direct Verification of Indentor Extension Types A, B, C, D, DO, O, OO  | 2.44 mm to 2.54 mm (0.096 in to 0.1 in)                                    | 0.0153 2 mm (0.000 603 in)   | Gage Blocks  |
| Durometers <sup>FO</sup><br>Indentor Spring Force Types A, B, C, D, DO, O, OO  | 0 kg to 4.53 kg  | 2.0 g  | Electronic Force gage                              |



# Certificate of Accreditation: Supplement

## Wayac Scales & Calibration, Inc.

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

Accreditation is granted to the facility to perform the following calibrations:

### Thermodynamics

| MEASURED INSTRUMENT, QUANTITY OR GAUGE                                     | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED                      |
|--|---|--|---|
| Thermometers <sup>FO</sup>   | 33 °C to 300 °C<br>(92 °F to 572 °F)        | 0.12 °C<br>(.22 °F)  | Fluke 2620A, Dry Block and PRT<br>Omega CL-27<br>Omega BB703 Black Body |
| Digital Temperature <sup>FO</sup> Controllers                              | 0 °C to 1 260 °C<br>(32 °F to 2 300 °F)     | 0.34 °C<br>(0.61 °F)   |   |
| Non-Contact Infrared Temperature Thermometers,<br>Pyrometers <sup>FO</sup> | -6.6 °C to 400 °C<br>(20.12 °F to 752 °F)   | 1.61 °C<br>(2.9 °F)  |   |
| Relative Humidity <sup>FO</sup>  | 15% to 95% RH<br>Non-condensing             | 4.9% RH  | Psychrometer  |
| Humidity Meters <sup>FO</sup>  | 5% and 95% RH                               | 2% RH  | Vaisala Salt Chamber  |
|  | 20 °C                                       | 0.52 °C  | PRT with Salts Chamber  |

### Time and Frequency

| MEASURED INSTRUMENT, QUANTITY OR GAUGE      | RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE | CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ ) | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED |
|---|---|--|--|
| Digital/Mechanical Tachometer <sup>FO</sup> | 40 rpm to 99 999 rpm                        | (1.18 + 0.003 % of reading) rpm  | Tachometer   |
| Stopwatches / Timers <sup>FO</sup>          | 60 s. to 24 hr                              | 0.5 s  | Universal Counter                                  |

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its Fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its Fixed location.





## *Certificate of Accreditation: Supplement*

### **Wayac Scales & Calibration, Inc.**

2899 Hilliard Rome Road, Hilliard, OH 43026

Tim Jarrell Phone: 614-529-4556

*Accreditation is granted to the facility to perform the following calibrations:*

4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its Fixed location and Onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its Fixed location and Onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories Fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The term R represents radius in inches or millimeters as appropriate to the uncertainty statement.

